

INSTRUCTIONS - GLOVE FOR ISOLATOR

MADE OF EPDM Black / EPDM White (Eb/Ew) 6/10mm MOUNTED ON SECURE CONNECTION RING (BCS)

➤ Scope/Characteristics of the material and expiry

- This glove can be used in the medical/pharmaceutical industries (domain of life sciences) and nuclear industries.
 - This glove is manufactured using EPDM black and EPDM White (Ethylene-Propylene-Diene Monomer)
 - This glove protects against certain chemical and mechanical risks, against radioactive contamination and provides protection against micro-organisms and viruses.
- It can be used in an atmosphere containing ozone. This glove does not protect against ionising radiation.
- The shelf life of this glove in its original packaging stored under the conditions described below is 3 years.

➤ Legislation, Risk analysis and recommendations for use


- This glove conforms to the regulation 2016/425.
- Please note, according to directive (UE) 89/656/CEE modified by directive 2019/1832, the necessity to carry out a risk analysis relating to wearing the glove (e.g. breaking of leak tightness).
- Before each use, it is the responsibility of the user to check the quality and the integrity of the glove. The glove should be discarded if it has tears, holes or changes in surface appearance or colour that can show an alteration due to chemical products.
- For the assembly and disassembly of the gloves on the gloveboxes, please follow the instructions drafted by the safety manager and the glove port manufacturer.
- In case of a contamination, follow the instructions drafted by the safety manager.
- The gloves should be used at ambient temperature. Please contact the manufacturer for use under other temperature conditions.
- The gloves must be worn on clean and dry hands, with the nails cut short. Avoid wearing jewellery.
- During the use of unspecified chemical products (those not mentioned in the list of chemical products), please contact the manufacturer for more information.
- This glove can only be mounted on a cuff equipped with the same device. The traction and integrity tests validated by IFTH have been conducted with an Eb/Ew (4/10 mm) glove and Eb (4/10 mm) cuff assembly. The choice of the cuff is made according to the risk analysis.
- Method of fastening between the glove and the cuff: The ring inseparable from the glove is assembled on the connected ring of the cuff by 3 pads mounted on flexible arms. The 3 pads are clipped on to the 3 grooves in order to guarantee the fastening.
- Method of fastening between the cuff and the containment enclosure: The cuff is fixed on a glove box glove port.
- The gloves must be worn, used in the initial side in which they were delivered by Piercan. In case of non-compliance, PIERCAN does not guarantee the performance claimed in the instruction manual.

➤ Components:

- This glove does not contain substances in proportions in which they are known or suspected to have harmful effects on the user's health or hygiene under the foreseeable conditions of use.


➤ Properties of the glove

- Protection against chemical products evaluated as per **EN ISO 374-1+ A1 :2018**
- Resistance to deterioration evaluated as per **EN374-4 :2013**

Tested products		Level of permeation (palm and cuff)	Average deterioration –palm (%)	Pictogram
A	Methanol	6 of 6	8.9	 ANP
N	Acetic acid 99%	Palm: 5 of 6	16.1	
P	Hydrogen peroxide 30%	6 of 6	12.1	

This information does not reflect the actual duration of protection at the workplace, or the differentiation between the mixtures and pure chemical products. The chemical resistance has been evaluated under laboratory conditions using the samples collected in the palm and the cuff and only concerns the tested chemical product. It can be different if it is used in a mixture. It is recommended to check that the gloves are suitable for the intended use because the conditions at the place of work can differ from the typical test conditions, depending on the temperature, abrasion and deterioration. When they are worn, the protective gloves provide less resistance to hazardous chemical products owing to the change in their physical properties. Movements, tears, friction or deterioration caused by contact with chemical products, etc. can considerably reduce the actual duration of use. For corrosive chemical products, deterioration can be the most important factor to be taken into account in the selection of chemical-resistant gloves. Before use, it is recommended to inspect the gloves in order to ensure that they have no defects or imperfections.

- Protection against viruses, bacteria and moulds evaluated as per **EN ISO 374-5: 2016**
- Tightness evaluated as per **EN 374-2: 2014**

Tests	Level	Pictogram
Airtightness	Compliant	 VIRUS
Water-tightness	Compliant	
Penetration by contaminated liquids under hydrostatic pressure (method B)	No passage	

The resistance to penetration has been evaluated under laboratory conditions and only pertain to the tested test specimen.

Attestations d'Examen UE de Type (AET) (EU type examination certifications) delivered by: I.F.T.H. (no. 0072) Avenue Guy de Collongue – 69134 ECULLY Cedex.

Monitoring organisation: AFNOR CERTIFICATION (no. 0333):11, rue Francis Pressensé-93571 LA PLAINE ST DENIS Cedex

PIERCAN: Impasse des Macareux ZI Huppain / 14520 PORT EN BESSIN FRANCE Tel.: 33 (0)2 31 21 73 80 Fax: 33 (0) 2 31 21 40 23 Email: piercan @ piercan.fr Website: www. Piercan.fr

Instructions: Edition of 30/06/2025 Rev 2



INSTRUCTIONS - GLOVE FOR ISOLATOR MADE OF EPDM Black / EPDM White (Eb/Ew) 6/10mm MOUNTED ON SECURE CONNECTION RING (BCS)

- Protection against mechanical risks evaluated as per **EN388+A1: 2018**

Tests	Level	Pictogram
Abrasion	3 of 4	
Cut by slicing	X of 5	
Tear	X of 4	
Perforation	1 of 4	
X: test not carried out		

The material consists of two materials (bi-layers). As a result, the overall rating does not necessarily reflect the performance of the outermost layer

- Precision, Integrity and ozone cracking


Tests	Level
Precision EN ISO 21420 + A1 :2024	3 of 5
Integrity (pressure 30 mbar) EN421-2010	Compliant
Ozone cracking resistance EN421-2010	4 of 4

The highest figures correspond to the highest performance levels.

➤ Marking

- Example of marking put on the glove and the packaging (see below)

LOGO PIERCAN



PIERCAN
(1)




XXXXXXXXXX
(2)

E10330 Eb/Ew 9 E6 2 BCS [XXXXX]

(4) (5) (6) (7)(8) (9) (10) (11)

FR XXXXXX LOT : XXXX **AAAA/MM**
(12) (13)



No.	DESCRIPTION
1	Manufacturer – In charge of release to the market
2	QR code and Unique IDentifier
3	“CE” marking and No. of the body in charge of the annual inspection
4	Form
5	Length in mm
6	Material
7	Size
8	Thickness in 10th of mm
9	Diameter of rim in mm
10	Glove mounted on secure connection ring
11	Product Item Code
12	Case number and batch number (“FR” French manufacturing “US” American manufacturing)
13	Date of shelf life in storage + pictogram
14	“INFORMATION” pictogram
15	“Protection against radioactive contamination” pictogram EN 421 :2010
16	“Protection against mechanical risks” pictogram EN 388+A1 :2018
17	“Protection against chemical risks” pictogram EN ISO 374-1 +A1 :2018
18	“Protection against viruses” pictogram EN ISO 374-5 :2016

➤ Size available in the main ambidextrous model

REFERENCE	Ø RDG (diamètre du rond de gant)	TAILLES
E10330	110	De 6 à 10

➤ Suitable packaging for transport

- The gloves should be transported in their original packaging.

➤ Storage

It is recommended to store gloves:

- In their original packaging, flat, black side of bag up and in their original box
- At a storage temperature between 5 to 35°C /41 to 95°F
- In a dry place away from direct light
- Away from electrical power source to prevent accelerated ageing.

Attestations d'Examen UE de Type (AET) (EU type examination certifications) delivered by: I.F.T.H. (no. 0072) Avenue Guy de Collongue – 69134 ECULLY Cedex.

Monitoring organisation: AFNOR CERTIFICATION (no. 0333):11, rue Francis Pressensé-93571 LA PLAINE ST DENIS Cedex

PIERCAN: Impasse des Macareux ZI Huppain / 14520 PORT EN BESSIN FRANCE Tel.: 33 (0)2 31 21 73 80 Fax: 33 (0) 2 31 21 40 23 Email: piercan @ piercan.fr Website: www. Piercan.fr

Instructions: Edition of 30/06/2025 Rev 2



INSTRUCTIONS - GLOVE FOR ISOLATOR MADE OF EPDM Black / EPDM White (Eb/Ew) 6/10mm MOUNTED ON SECURE CONNECTION RING (BCS)

➤ Cleaning/decontamination and maintenance

- Gloves are not designed to be laundered.
- Prior to the use of unspecified chemicals, please contact the manufacturer for compatibility information.
- Do not contact gloves with sharp or pointed objects such as wire brushes, sandpaper, screwdrivers, or similar objects.
- If contaminated with chemicals, the gloves are for single use only

➤ Treatment of the glove

- Non-recyclable gloves
- It can be discarded as simple non-toxic waste if it is not soiled by a hazardous product
- If a glove is soiled, discard it through the appropriate channel

The EU Declaration of Conformity is available at: www.piercan.fr